

SAMSUNG ELECTRONICS Co., Ltd.

510(k) Premarket Notification - Traditional

SAMSUNG

K23105
Page 1 of 4

510(k) Summary

JUN 07 2013

This summary of 510(k) safety and effectiveness information is being submitted accordance with requirements of 21 CFR 807.92

1. Date: Sep. 20, 2012

2. Submitter

- A. Company Name: SAMSUNG ELECTRONICS Co., Ltd.
- B. Address: 129, Samsung-ro, Yeongtong-gu, Suwon-si, Gyeonggi-do, 443-742, Republic of Korea

3. Primary Contact Person

- A. Name: Charlie Mack
- B. Title: President of International Regulatory Consultants
- C. Phone Number: 541-376-5063
- D. FAX Number: 541-376-5063
- E. E-Mail: charliemack@irc-us.com

4. Secondary Contact Person

- A. Name: Shim, Kyungam
- B. Title: Regulatory Affairs Manager
- C. Phone Number: +82-31-200-3356
- D. FAX Number: +82-31-200-0700
- E. E-Mail: kyungam.shim@samsung.com

5. Device

- A. Trade Name: XGEO GU60
- B. Common Name: Digital Diagnostic X-ray System
- C. Classification Name: System, X-ray, Stationary
- D. Product Code: KPR

6. Predicate Device

- A. Manufacturer: DRGEM Corporation
- B. Trade Name: Diamond-5A
- C. 510(k) Number: K102408



510(k) Premarket Notification - Traditional

D. 510(k) Decision Date: 2010 Dec 27

A. Manufacturer: General Electric Company

B. Trade Name: Revolution XR/d Digital Radiographic Imaging System

C. 510(k) Number: K012389

D. 510(k) Decision Date: 2001 Aug 10

7. Device Description

The XGEO GU60 digital X-ray imaging system is to be used to take and store image for diagnosis of patients. It consists of the High voltage generator (HVG), U-arm positioner, Detector, X-ray tube, Collimator and etc.

8. Intended Use

The XGEO GU60 Digital X-ray Imaging System is intended for use in generating radiographic images of human anatomy by a qualified/trained doctor or technician. This device is not intended for mammographic applications.

9. Comparison with predicate device :

Samsung Electronics Co., Ltd., believes that the XGEO GU60 is substantially equivalent to the Revolution XR/d of General Electric Company and Diamond-5A of DRGEM Corp.

Comparison with the predicate device made through the system component items such as detector, generator etc., has been proven to be similar in many ways, but differences in the seven items were found. Specifically, even if the differences of the capacity and size of High Voltage Generator, Ceiling Suspension, Wall stand and Patient table exists, these are considered minor impact on the safety and performance. Also, Collimator, Detector, and Image Process Function in terms of the design and the technology characteristic have differences in the following characteristics:

1) Blade moving method: Motorized method for adjusting a collimator radiation field is only available. It just concerns on the method for control a radiation field and has not effect on safety.

2) Detector type : Phosphor of detector is different. Two major scintillation materials such as Gd2O2S and CsI are commonly used. It affects the image quality but it has not effect on safety.

3) Number of pixels : Number of pixels is different. A detector with a large number of



510(k) Premarket Notification - Traditional

pixels can cover a wider area of body. It has not effect on safety. (Difference from Revolution XR/d Digital Radiographic Imaging System of General Electric Company, but Similarities to DIAMOND-5A of DRGEM Corporation)

4) Pixel pitch : Pixel pitch, a measurement that indicates the distance between pixels and determines an image resolution, is different. It affects the image quality but it has not effect on safety.

5) High contrast limiting resolution : High contrast limiting resolution is different due to the difference of the pixel pitch. The resolution is higher as the value of the pixel pitch is lower. It affects the image quality but it has not effect on safety.

6) DQE : The value of DQE is different. The detector type with CSI phosphor has higher DQE value. It affects the image quality but it has not effect on safety.

7) MTF : The value of MTF is different. The detector type with CSI phosphor has higher MTF value. It affects the image quality but it has not effect on safety.

However, these differences do not have an effect on safety and efficiency compared with the predicate device, Revolution XR/d of General Electric Company and Diamond-5A of DRGEM Corp..

In summary, the XGEO GU60 does not introduce any new potential safety risks and is substantially equivalent to and performs as well as the predicate device.

In conclusion, the XGEO GU60 is substantially equivalent to Revolution XR/d of General Electric Company and Diamond-5A of DRGEM Corp..

10. Safety, EMC and Performance Data :

Electrical, mechanical, environmental safety and performance testing according to standard IEC 60601-1, IEC 60601-1-1, IEC 60601-1-3, IEC 60601-2-7, IEC 60601-2-28, IEC 60601-2-32, IEC 60601-2-54, 21CFR1020.30 and 21CFR1020.31 were performed, and EMC testing was conducted in accordance with standard IEC 60601-1-2(2001). And Non-clinical testing and Clinical Testing were conducted in accordance with Solid-state X-ray Imaging System Guidance. All test results were satisfied.

11. Conclusions :

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the information provided in this premarket notification Samsung Electronics Co., Ltd. concludes that The XGEO GU60 is safe and effective and substantially

K123105
Page 4 of 4

SAMSUNG ELECTRONICS Co., Ltd.

SAMSUNG

510(k) Premarket Notification - Traditional
equivalent to predicate devices as described herein.

12. Samsung Electronics Co., Ltd. will update and include in this summary any other information deemed seasonably necessary by the FDA.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

June 7, 2013

SAMSUNG ELECTRONICS Co., Ltd.
% Mr. Charlie Mack
Principal Engineer
International Regulatory Consultants
77325 Joyce Way
ECHO OR 97826

Re: K123105

Trade/Device Name: XGEO GU60
Regulation Number: 21 CFR 892.1680
Regulation Name: Stationary x-ray system
Regulatory Class: II
Product Code: KPR
Dated: May 21, 2013
Received: May 29, 2013

Dear Mr. Mack:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for

Janine M. Morris
Director, Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K123105

Device Name: XGEO GU60

Indications for Use:

The XGEO GU60 Digital X-ray Imaging System is intended for use in generating radiographic images of human anatomy by a qualified/trained doctor or technician. This device is not intended for mammographic applications.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of *In Vitro* Diagnostics and Radiological Health (OIR)



(Division Sign-Off)
Division of Radiological Health
Office of *In Vitro* Diagnostics and Radiological Health

510(k) K123105